



CERTIFICATE OF COMPLIANCE

Certificate Number: SGSNA/17/SUW/00122

Contract Number: 710292
Certificate Project Number: SUW-CERT170500144

Certified Product: RT300 System
Trademarks: Restorative Therapies
Model(s): RT300 Supine Cycle Ergometer
Technical Data: 120V, 60 Hz, 432VA; (120V version)
120/230V, 50/60Hz, 432VA (120/230V version)
Class I, mobile, Type BF applied parts

Certificate Holder: Restorative Therapies, Inc.
1434 Fleet Street, 21231, Baltimore, Maryland, United States of America

This certificate supercedes previous certificates issued with the same certificate number. Certification is valid when products are indicated on the SGS directory of certified products at www.sgs.com or using the QR code below. The product is certified according to ISO/IEC Guide 17067, Conformity assessment - Fundamentals of product certification, System 3, and in accordance with:

ANSI/AAMI ES60601-1:2005 +C1 (2009) +A1 (2012) +A2 (2010)
CAN/CSA C22.2 No. 60601-1:14
CAN/CSA C22.2 No. 60601-2-10:14
IEC 60601-2-10:2012 + A1:2016

Authorized by:

Effective date: 19 June 2020



Paul Krauss
Certifier



Certification Body

Consumer and Retail Services, a division of SGS North America Inc.
620 Old Peachtree Road, Ste. 100, Suwanee, GA 30024, USA
t +1 770 570 1800 f +1 770 277 1240 www.sgs.com

