

Contract Number: 710292

Certificate: **SGSNA/17/SUW/00122**
Test Report Number: 617046/01

Certified Product: RT300 System
Trademark (if any): Restorative Therapies
Model(s): RT300 Supine Cycle Ergometer
Technical Data: 120V, 60 Hz, 432VA; Class I, mobile, Type BF applied parts



Name and Address of Applicant: Restorative Therapies, Inc.
1434 Fleet Street, 21231, Baltimore, Maryland, United States of America

Name and Address of Manufacturer: Restorative Therapies, Inc.
1434 Fleet Street, 21231, Baltimore, Maryland, United States of America

Name and Address of Factory: Restorative Therapies, Inc.
1434 Fleet Street, 21231, Baltimore, Maryland, United States of America

This certificate supercedes previous certificates issued with the same certificate number. Certification is valid when products are indicated on the SGS directory of certified products at www.sgs.com. The product is certified according to ISO/IEC Guide 17067, *Conformity assessment - Fundamentals of product certification, System 3*, and in accordance with

ANSI/AAMI ES60601-1:2005 +C1 (2009) +A1 (2012) +A2 (2010)
CAN/CSA C22.2 No. 60601-1:14
CAN/CSA C22.2 No. 60601-2-10:14
IEC 60601-2-10

Authorized by

Effective date: 24 May 2017


Paul Krauss
Certifier

Consumer and Retail Services, a division of SGS North America Inc.
620 Old Peachtree Road, Ste. 100, Suwanee, GA 30024, USA
t +1 770 570 1800 f +1 770 277 1240 www.sgs.com
Page 1 of 1



This certificate is issued by the company under its General Conditions for Certification Services accessible at www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitations of liability defined therein and in the Test Report here above mentioned which findings are reflected in this Certificate. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.